

## Konformitätserklärung

Certificate of Compliance

### ERWEKA Release Rate Tester, Typ RRT 10

ERWEKA Release Rate Tester, Type RRT 10



Für den ERWEKA Release Rate Tester, Typ RRT 10, bescheinigen wir in alleiniger Verantwortung die Übereinstimmung mit den Forderungen der European Pharmacopoeia 9<sup>th</sup> Edition, Kapitel 2.9.3, der USP40, General Chapter <711>, Apparatus 3 und/oder der USP40, General Chapter <724>, Apparatus 7<sup>o</sup>.

Hinweis: Eintauchende Zylinder, Glasbehälter, Haltestäbe etc. verfügen jeweils über separate Konformitätserklärungen.

For the ERWEKA Release Rate Tester, Type RRT 10, we declare in our sole responsibility the conformity to European Pharmacopoeia 9<sup>th</sup> Edition, chapter 2.9.3, USP40, General Chapter <711>, Apparatus 3 and/or the USP40, General Chapter <724>, Apparatus 7<sup>o</sup>.

Note: Declaration of conformities regarding reciprocating cylinder, vessels and reciprocating holder etc. are issued separately.

Langen, 2019-05-13

i.A. S. Klier  
Sevgi Klier

- Head of Integrated Management System -

① Je nach Geräteausführung  
Depending on device version