

Konformitätserklärung

Certificate of Compliance

ERWEKA Release Rate Tester, Typ RRT 10

ERWEKA Release Rate Tester, Type RRT 10



Für den ERWEKA Release Rate Tester, Typ RRT 10, bescheinigen wir in alleiniger Verantwortung die Übereinstimmung mit den Forderungen der European Pharmacopoeia 10th Edition, Kapitel 2.9.3, der USP42, General Chapter <711>, Apparatus 3 und/oder der USP42, General Chapter <724>, Apparatus 7[Ⓞ].

Hinweis: Eintauchende Zylinder, Glasbehälter, Haltestäbe etc. verfügen jeweils über separate Konformitätserklärungen.

For the ERWEKA Release Rate Tester, Type RRT 10, we declare in our sole responsibility the conformity to European Pharmacopoeia 10th Edition, chapter 2.9.3, USP42, General Chapter <711>, Apparatus 3 and/or the USP42, General Chapter <724>, Apparatus 7[Ⓞ].

Note: Declaration of conformities regarding reciprocating cylinder, vessels and reciprocating holder etc. are issued separately.

Langen, 2020-07-03

i.A.


Günter Annacker
- Quality department -

Ⓞ Je nach Geräteausführung
Depending on device version